PDS No. 1262xx, 1262xx-2D1, 1262xx-2DG	PRODUCT DATA SHEET	Page 1 of 2				
Revision 10	Cryo.s™, 2 ml, Round Bottom, Starfoot Base					
Revision to	Item-No. 1262xx, 1262xx-2D1, 1262xx-2DG	greiner				
Valid for Item-No.:	126261 (sterile) 126263 (sterile) 126277 (sterile) 12	.6278 (sterile)				
	126279 (sterile) 126280 (sterile) 1262xx-2D1 (sterile) 12	262xx-2DG (sterile)				

1.	Description / Specification			
1.1	Description	Cryo.s™, 2 ml, with different coloured screw caps, external thread, round bottom, starfoot base and additional cap inserts for labelling, sterile		
		126261: without writing area		
		126263, -277, -278, -279, -280: with writing area and graduation		
		1262xx-2DG: with linear barcode 128, human readable ID, graduation and ECC200 Datamatrix-Code, symbol size 14x14, on tube bottom, code sequence 'G600000001', 'G600000002' etc., for screw cap colour see 1.5		
		1262xx-2D1: with linear barcode 128, human readable ID, graduation and ECC200 Datamatrix-Code, symbol size 14x14, on tube bottom, code sequence customized as specified in order form CatNo. F071005, for screw cap colour see 1.5		
		Additional inserts, CatNo. 304 134 (50 pieces per bag)		
1.2	Dimensions	See Customer Drawing		
1.3	Volume	Working volume: up to 2.0 ml		
1.4	Material / Resin	Tube: PP (Polypropylene), free of heavy metal Screw cap: HDPE (High Density Polyethylene), free of heavy metal Small cap insert: PS (Polystyrene), free of heavy metal 126263-2DG, 1262xx-2D1: Datamatrix disc: PP (Polypropylene)		
1.5	Colour	Tube: natural, writing area and graduation: white		
		<u>Screw cap</u> : 126261, -263: natural		
		126277: green		
		126278: yellow		
		126279: blue 126280: red		
		1262xx-2D1, 1262xx-2DG: xx = colour code of screw cap with		
		xx = 63 for natural screw cap, $xx = 77$ for green screw cap,		
		xx = 65 for hatural screw cap, xx = 77 for green screw cap, xx = 78 for yellow screw cap, xx = 79 for blue screw cap,		
		xx = 80 for red screw cap, $xx = 79$ for blue screw cap,		
		Small cap insert: white		
		126263-2D1, 1262xx-2DG: white Datamatrix on black disc		
1.6	Sterilization	SAL 10 ⁻⁶		
1.7	Quality Control	Raw Material-Control: physical testing		
		Product-Control: testing of attributive and variable characteristics in		
		accordance with the valid specification		
1.8	Other Information	For single use only		

Data Sheet subject to change without notice!

Prior Issue	Drawn	Approved	Released	CONFIDENTIAL: Information contained in this
Revision	Date	Date	Date	document or drawing is confidential and proprietory to Greiner Bio-One GmbH. This
09	14 May 2020	14 May 2020	15 May 2020	document may not be reproduced for any
Date	Name	Name	Name	reason without written permission from Greiner Bio-One GmbH. All rights of design, invention,
17.12.2019	S. Kaelberer	Dr. S. Mühlfriedel	A. Illig	and copyright are reserved.

PDS No. 1262xx, 1262xx-2D1, 1262xx-2DG	PF	RODUCT DATA S	SHEET		Page 2 of 2
Revision 10	Cryo.s™, 2 ml	, Round Bottom,	Starfoot Base C	ϵ	9
Revision 10	Item-No.	1262xx, 1262xx-2D	1, 1262xx-2DG		greiner
Valid for Item-No.:	126261 (sterile)	126263 (sterile)	126277 (sterile)	12	26278 (sterile)
	126279 (sterile)	126280 (sterile)	1262xx-2D1 (sterile)	12	262xx-2DG (sterile)

2.	Features	
2.1	Basic features	Free of detectable DNase/RNase, human DNA and pyrogens.
		Contents non-cytotoxic
2.2.	Temperature range	-196°C to +100°C
2.3	Autoclavability	No
2.4	Centrifugation, max. RCF	5800 x g: swinging-bucket rotor
		26000 x g: fixed-angle rotor
2.5	Chemical Resistance	See homepage:
		https://www.gbo.com/en_INT/know-how-services/download-center.html
2.6	Shelf life	5 years
2.7	Other Information	Recommended for storage in the freezer or above liquid nitrogen.
		Must not be used in liquid phase of nitrogen!

3.	Packaging		
3.1	Pieces / Bag	100	
3.2	Pieces / Box	500	
3.3	Lot-No.	E YY MM XXX (manufacturing facility, year, month, consecutive SAP-No.)	
3.4	Other Information	Additional in each box: Safety precaution	
		3 bags with 50 cap inserts per bag (150 pieces)	
		Certificate of Quality to download	

4.	Other Information
4.1	Laboratory use only

Data Sheet subject to change without notice!

Prior Issue	Drawn	Approved	Released	CONFIDENTIAL: Information contained in this
Revision	Date	Date	Date	document or drawing is confidential and proprietory to Greiner Bio-One GmbH. This
09	14 May 2020	14 May 2020	15 May 2020	document may not be reproduced for any
Date	Name	Name	Name	reason without written permission from Greiner Bio-One GmbH. All rights of design, invention,
17.12.2019	S. Kaelberer	Dr. S. Mühlfriedel	A. Illig	and copyright are reserved.